

CYSTIC FIBROSIS (CF) MULTIPLEX REAL-TIME PCR KIT (18 MUTATIONS)

Cat. No: 14R-20-18

INTRODUCTION

Cystic fibrosis (CF) is one of the most common life-limiting hereditary disorders, primarily affecting the lungs, pancreas, and other organs through the production of thick and sticky mucus. It is caused by mutations in the cystic fibrosis transmembrane conductance regulator (CFTR) gene, which regulates the composition of sweat, digestive fluids, and mucus. The Cystic Fibrosis (CF) Multiplex Real-Time PCR Kit is designed to detect 18 clinically relevant mutations in the CFTR gene ⁽¹⁻⁴⁾.

INTENDED USE

Cystic Fibrosis (CF) Multiplex Real-Time PCR Kit can detect G542X, W1282X, E92K, 1677delTA, K68E, 2789+5G>A, ΔF508, 1717-1G>A, 4010delTATT, F1052V, I148T, G551D, Y1032C, 3849+10kbC>T, R560T, 2183AA--G, S4X, N1303K mutations of CFTR gene in whole blood samples by using qualitative Real-Time PCR method.

TARGETED USER

For professional use only. Testing should be performed by professionals trained in molecular techniques.

PRINCIPLE OF THE SYSTEM

During the PCR reaction, the DNA polymerase cleaves the probe at the 5' end and separates the reporter dye from the quencher dye only when the probe hybridizes perfectly to the target DNA. This cleavage results in the fluorescent signal which is monitored by Real-Time PCR detection system. An increase in the fluorescent signal (C_T) is proportional to the amount of the specific PCR product ^(5,6).

PRODUCT SPECIFICATION

Each isolated DNA should be tested with all master mixes separately. The kit provides reagents in a "ready-to-use" master mix format which has been specifically adapted to 5' nuclease PCR for SNP analysis. The test system is designed by SNP Biotechnology for use with sequence specific primers and probes.

The fluorescence of mutation analysis is FAM, HEX/JOE and Texas RED. Also each master mix contains an internal control labelled with CY5 dye (See Table 2). Internal Control is Prothrombin gene – FII (OMIM: 176930).

SYSTEM CONTENTS

Reagents	10 rxns	20 rxns	50 rxns
CF Mix 1	200 µl	400 µl	1000 µl
CF Mix 2	200 µl	400 µl	1000 µl
CF Mix 3	200 µl	400 µl	1000 µl
CF Mix 4	200 µl	400 µl	1000 µl
CF Mix 5	200 µl	400 µl	1000 µl
CF Mix 6	200 µl	400 µl	1000 µl
CF Mix 7	200 µl	400 µl	1000 µl
CF Mix 8	200 µl	400 µl	1000 µl
CF Mix 9	200 µl	400 µl	1000 µl
CF Mix 10	200 µl	400 µl	1000 µl
CF Mix 11	200 µl	400 µl	1000 µl
CF Mix 12	200 µl	400 µl	1000 µl
Control DNA*	100 µl	150 µl	300 µl

Table 1: Kit content

* Control DNA is a synthetic plasmid containing Wild Type for all mutations/polymorphisms of CFTR gene. Since to Control DNA is a synthetic plasmid, it does not amplify CY5 dye and amplification plots of control DNA may appear slightly different from the sample DNA. Please gently vortex and then spin centrifuge for 1-2 seconds before using the Control DNA.

STORAGE

- All reagents should be stored at – 20 °C and dark.
- All reagents can be used until the expiration date on the box label.
- Repeated thawing and freezing (>4X) should be avoided, as this may reduce the sensitivity of the assay.

MUTATION DYE TABLE

Tubes	Mutations	Dyes
Mix 1	G542X Wild Type	FAM
	W1282X Wild Type	HEX / JOE
	E92K Wild Type	Texas Red
	Internal Control	CY5
Mix 2	G542X Mutant	FAM
	W1282X Mutant	HEX / JOE
	E92K Mutant	Texas Red
	Internal Control	CY5
Mix 3	1677delTA Wild Type	FAM
	K68E Wild Type	HEX / JOE
	2789+5G>A Wild Type	Texas Red
	Internal Control	CY5
Mix 4	1677delTA Mutant	FAM
	K68E Mutant	HEX / JOE
	2789+5G>A Mutant	Texas Red
	Internal Control	CY5
Mix 5	ΔF508 Wild Type	FAM
	1717-1G>A Wild Type	HEX / JOE
	4010delTATT Wild Type	Texas Red
	Internal Control	CY5
Mix 6	ΔF508 Mutant	FAM
	1717-1G>A Mutant	HEX / JOE
	4010delTATT Mutant	Texas Red
	Internal Control	CY5
Mix 7	F1052V Wild Type	FAM
	I148T Wild Type	HEX / JOE
	G551D Wild Type	Texas Red
	Internal Control	CY5
Mix 8	F1052V Mutant	FAM
	I148T Mutant	HEX / JOE
	G551D Mutant	Texas Red
	Internal Control	CY5
Mix 9	Y1032C Wild Type	FAM
	3849+10kbC>T Wild Type	HEX / JOE
	R560T Wild Type	Texas Red
	Internal Control	CY5
Mix 10	Y1032C Mutant	FAM
	3849+10kbC>T Mutant	HEX / JOE
	R560T Mutant	Texas Red
	Internal Control	CY5
Mix 11	2183AA--G Wild Type	FAM
	S4X Wild Type	HEX / JOE
	N1303K Wild Type	Texas Red
	Internal Control	CY5
Mix 12	2183AA--G Mutant	FAM
	S4X Mutant	HEX / JOE
	N1303K Mutant	Texas Red
	Internal Control	CY5

Table 2: Tubes - mutations - dyes.

SAMPLE COLLECTION

Cystic Fibrosis (CF) Multiplex Real-Time PCR Kit is approved for use with whole blood samples.

- Standard precautionary instructions must be followed by all healthcare professionals during the collection and transportation of whole blood samples.

- Whole blood samples should be collected in appropriate containers before delivery to the laboratory.

- Freezing and thawing of samples should be avoided.

DNA EXTRACTION

Blood samples should be collected in appropriate sterile EDTA tubes and can be stored at +4°C up to one month. For more than one month specimen should be stored at -20°C. It is advised to gently mix the tube (with EDTA) after collection of blood to avoid coagulation.

Our system optimized according to GeneAII® Exgene™ Blood SV. It is advised to elute DNA with 150 µl elution buffer for better results.

PROCEDURE

- Different test tubes should be prepared for each master mix.
- Leave the master mixes* and controls at RT to melt.
- Before starting work, mix the master mixes gently by pipetting
- For each sample, pipet **20 µl master mix** with micropipets of sterile filter tips to each optical white strips or tubes.
- Add **5 µl DNA** into each tube. Please do not pipette DNA before and after addition into well.
- Optical caps are closed, it is recommended to spin the plates/strips at low speed for a short time.
- Run with the programme shown below.

*Master mixes include HotStart Taq DNA Polymerase.

PCR PROGRAMME

95 °C	5 Sec.	32 Cycles
60 °C	45 Sec.	

Table 3: PCR Programme

Fluorescent dyes are FAM, HEX/JOE, Texas Red and CY5.

This system can be used with the following devices;

- Bio-Rad CFX96, Opus 96
- ABI Prism® 7500/7500 Fast
- Mic qPCR Cycler

For other four or more channel Real-Time PCR devices (which can read FAM, HEX/JOE, Texas Red and CY5 dyes), a trial run is recommended.

If you use:

ABI Prism® system, please choose **"none"** as passive reference and quencher.

Mic qPCR Cycler, please adjust gain settings, **"Green Auto Gain"** to **20** and **"Yellow Auto Gain"** to **10**.

Supplied Materials

- White PCR plates/strips with optical covers*

*The PCR Plate/strip tube and caps seriously affect the amplification curve quality. Therefore, white PCR plates/strips and optical caps provided by the manufacturer should be used with the kit.

Required Materials (Not Provided)

- PCR Cabinet
- Vortex Mixer
- Desktop Microcentrifuge (For 2.0ml tubes and PCR strip tubes), plate spin for studies using PCR plates.
- Automated or spin column based DNA isolation Kit
- Disposable powder-free laboratory gloves
- Micropipettes (0.5ml-1000ml)
- Micropipette tips
- Standard laboratory equipments.

DATA ANALYSIS

After the run is completed data are analysed using the software with FAM, HEX/JOE, Texas Red and CY5 dyes. The below results were studied with Bio-Rad CFX96. The threshold values for all dyes were set to 500, based on experiments conducted using the Bio-Rad CFX96 Real-Time PCR system, the GeneAII® Exgene™ Blood SV Isolation Kit, and white PCR strips supplied by SNP Biotechnology. Threshold values may vary depending on the PCR device, DNA isolation kit, and the type or brand of PCR strips/tubes used.

Internal control amplification plots must be seen in all wells except NTC and has been labelled with CY5 dye. The C_T value of internal controls should be $21 \leq C_T \leq 28$. These values are optimised according to the GeneAII® Exgene™ Blood SV Isolation Kit and Bio-Rad CFX96 Real-Time PCR Device. C_T values may vary $\pm 2/3$ cycle according to the other DNA isolation systems and Real-Time PCR devices.

Amplification plots of mutations can be analysed by FAM, HEX/JOE and Texas Red dyes. The C_T value should be between $21 \leq C_T \leq 28$. These values are optimised according to the GeneAII® Exgene™ Blood SV Isolation Kit and Bio-Rad CFX96 Real-Time PCR Device. C_T values may vary $\pm 2/3$ cycle according to the other DNA isolation systems and Real-Time PCR devices.

- Homozygous wild type sample gives amplification signal only with wild type master mix.
- Heterozygous sample gives amplification signal both with wild type and mutant master mixes.
- Homozygous mutant sample gives amplification signal only with mutant master mix.
- The difference of the C_T value with wild type and mutant amplification plots should be ≤ 3 for heterozygote sample. If it is $4 \leq C_T \leq 6$, test should be repeated, if >6 , the late plot should be considered as non-specific.




Figure 1: Internal Control plots – CY5 Dye




Figure 2: G542X Homozygous Mutant Sample (FAM Dye)




Figure 3: I148T Heterozygous Sample (HEX/JOE Dye)




Figure 4: N1303K Homozygous Wild Type Sample (Texas Red)

CAUTIONS

- All reagents should be stored at suitable conditions.
- Do not use the PCR master mixes forgotten at room temperature.
- Thaw PCR master mix at room temperature and slowly mix by inverting before use.
- Shelf-life of PCR master mix is 12 months. Please check the manufacturing data before use.
- Only use in vitro diagnostics.

DISPOSAL OF KIT

Dispose of it according to the legal regulations of your region

SYMBOLS AND DESCRIPTIONS












	Catalog Number		CE Mark
	Lot Number		Unique Device Identifier (01)Device Identifier (17)Expiry Date (10)Lot Number
	Manufacturer		Test Quantity
	Fragile		Storage Temperature
	Protect from directly sunlight		In Vitro Diagnostics
	Expiry Date		

Table 4: Symbols and descriptions

REFERENCES

1. John E. Mickle And Garry R. Cutting. "Genotype-Phenotype Relationships In Cystic Fibrosis". Medical Clinics Of North America. May 2000. Volume 84 - Number 3.
2. Chantal Farra, RitaMedawar, SalmanMroueh, Myrna Souaid, Faiza Cabet and JohnnyAwwad. "Cystic fibrosis: A new mutation in the Lebanese population". Journal of Cystic Fibrosis 7 (2008) 429-432.
3. Phillis Lakeman, Johan J.P. Gille, Jeannette E. Dankert-Roelse, Harry G.M. Heijerman, Anne Munck, Albert Iron, Hartmut Grasmann, Antje Schuster, Martina C. Cornel and Leo P. ten Kate. "CFTR Mutations in Turkish and North African Cystic Fibrosis Patients in Europe: Implications for Screening". Genetic Testing · April 2008. Volume 12, Number 1.
4. N.V. Petrova, N.Yu Kashirskaya, T.A. Vasilyeva, E.E. Timkovskaya, A.Yu Voronkova, L.A. Shabalova, E.I. Kondratyeva, V.D. Sherman, O.G. Novoselova, N.I. Kapranov, R.A. Zinchenko, E.K. Ginter, A.Kh-M. Makaov and B. Kerem. "High prevalence of W1282x mutation in cystic fibrosis patients from Karachay-Cherkessia". Journal of Cystic Fibrosis 15 (2016) e28-e32
5. Yolanda S Lie and Christos J Petropoulos. "Advances in quantitative PCR technology: 5' nuclease assays". Current Opinion in Biotechnology Volume 9, Issue 1 February 1998, Pages 43-48.
6. Luis Ugozzoli and R. Bruce Wallace. "Allele-Specific Polymerase Chain Reaction". A Companion to Methods in Enzymology Vol. 2, No. 1, February, pp. 42-48, 1991.

TROUBLESHOOTING PROBLEMS AND SOLUTIONS

Problem	Reason	Solution
Internal control does not work/ low amplification	Absence of DNA / DNA extraction problems	Repeat test
	Absence of DNA / DNA extraction problems	• DNA extraction should be repeated.
	Sample is containing PCR inhibitor(s)	• DNA extraction should be replaced with one of the recommended methods.
No target gene amplification curves in some samples for both wild type and mutant mixes.	Absence of DNA / not added into well	Repeat test
	Absence of DNA / DNA extraction problems	• DNA extraction should be repeated.
	Sample is containing PCR inhibitor(s)	• DNA extraction should be replaced with one of the recommended methods.
No target DNA/internal control amplification curves in all wells	Error in temperature/time settings in PCR program	Correct any errors in the temperature/time settings in the PCR Program and repeat the test.
	Sample is containing PCR inhibitor(s)	• DNA extraction should be repeated. • DNA extraction should be replaced with one of the recommended methods.
Positive control result and/or C _T values are lower or higher than the value mentioned in User Manual.	Error in temperature/time settings in PCR program	Correct any errors in the temperature/time settings in the PCR Program and repeat the test.
C _T values are not valid (higher or lower) according to User Manual	Excessive or insufficient DNA sample	• Repeat the test. • DNA extraction should be repeated.
Low and/or invalid amplification curves	Stability problems arising from repeated thawing and freezing (>4X)	Repeated thawing and freezing (>4X) should be avoided, as this may reduce the sensitivity of the assay.
	Sample is containing PCR inhibitor(s)	• DNA extraction should be repeated. • DNA extraction should be replaced with one of the recommended methods.
	Stability problems arising from unavailable storage conditions.	All reagents should be stored at – 20 °C and dark.
	Bubble formation or pipetting error during pipetting	After adding the master mix and sample, it is recommended to spin the plates/strips at low speed for a short time.
For further questions, please contact us tech@snp.com.tr		

Table 5: Troubleshooting problems and solutions